



einstein™
imagine • explore • learn

Dissolved Oxygen Sensor with Temperature Compensation

Product Number: ENOXT422A



Overview

The Oxygen sensor is a maintenance-free galvanic oxygen electrode, capable of measuring Oxygen in solutions. With built-in temperature compensation, this sensor provides an easy way for students to receive accurate results without having to perform complicated compensation calculations.

The Oxygen sensor can be used to perform a wide variety of experiments taking place in a number of different environments such as measuring the amount of oxygen in an aquarium or understanding the effects of germination on oxygen levels.

The Oxygen Sensor can be connected to all types of einstein™ data loggers.

Typical experiments



Biology

- Monitoring changes in oxygen levels during photosynthesis and respiration of plants
- Consumption of oxygen by yeast during respiration of sugars
- Fermentation of yogurt



Water Quality

- Monitoring dissolved oxygen concentrations resulting from photosynthesis and respiration in an aquarium containing plants and/or fish.
- Monitoring the pattern between light and dissolved oxygen levels, in an aquarium with pondweed.
- On-site testing in streams and ponds to evaluate the capability of their water to support different types of plant and animal life.



Chemistry

- Oxidation of metals

Sensor specification

Range:	0 to 14 mg/L DO
Accuracy:	± 2% over entire range (after calibration)
Resolution (12-bit):	0.003 mg/L
Temperature Compensation:	Yes
Temperature Compensation range:	0 to 40 °C
Response Time for 95% of Reading:	Up to 2 minutes
Default Sample Rate	10 samples per second

Note: sensor cables sold separately

Technical Notes

- Before use, let the sensor adjust to the surrounding temp for several minutes.
- Before use, remove the protective cap and use the dropper to fill it with Galvanic O₂ fill solution. Make sure it contains no air bubbles. This procedure is necessary only if you have had to replace the electrode membrane or the Galvanic O₂ fill solution.
- It is important to stir the solution during measurement as oxygen consumption by the probe can momentarily reduce the oxygen concentration at the probe surface.
- Do not allow any air bubbles to be trapped on the membrane surface, as it will read an air bubble as in oxygen-saturated sample.
- The DO electrode membrane is very sturdy, relative to other electrodes. Its function is more of a physical barrier (permeable to O₂) rather than a chemical reaction type membrane. In pond water, its useful life is determined by how well it is cared for e.g., by avoiding stretching, scratching, and fouling the membrane.
- The membranes only problem areas are:
 - Scratches or tears - Scratches could change the accuracy of the probe, tears will cause immediate failure.
 - Fouling of the membrane - Impacts measurement only if it impedes the membranes permeability to DO. This is unlikely in pond water whereas in wastewater this may be a common problem.
 - Fouling of the Galvanic O₂ fill solution impacts measurement. If needed, replace the solution or the solution together with the membrane.

- Stretching of the membrane impacts measurement as the thickness of the membrane and position of the membrane on the cathode are critical parameters of the measurement.


Maintenance and Storage

After the electrode has been used for a period of time the following cleaning method is recommended:

1. Carefully unscrew the cap from the body.
2. Rinse the inside of the membrane cartridge with DI water.
3. Soak the lead anode in 0.1N HCL (8.3 mL concentrated HCL in 1000 mL DI water) for 15 to 20 minutes. Rinse the lead anode with DI water, blot dry with tissue paper. Screw the lead anode back onto the body.
4. Rinse the soaked portions thoroughly with DI water. Blot dry with tissue paper.
5. Fill the cap with Galvanic O₂ Fill Solution to a level just above the membrane cartridge using the dropper provided with the sensor. Hold the electrode in an upright position and gently screw the cap back onto the body.
6. Inspect the membrane for tears or leakage. The membrane should be uniformly stretched across the inner body. Replace the membrane if any damage has occurred.

Membrane Replacement

1. The membrane should be examined routinely after each fermentation cycle and replaced if any deterioration is evident. See the illustration below.
2. Carefully unscrew the cap from the body.
3. Using the membrane tool provided, pop the membrane cartridge from the cap.
4. Inspect the platinum cathode. Gently wipe it with tissue paper and see if there are any cracks or damage.
5. Take a new membrane out of the plastic package and insert into the cap. Push it down inside until firmly seated inside the outer cap using the membrane tool provided.
6. Fill the cap with electrolyte to a level just above the membrane cartridge, using the dropper provided with the sensor.
7. Do not allow any air bubbles to be trapped on the membrane surface, as it will read an air bubble as an oxygen-saturated sample.
8. Hold the electrode in an upright position and gently screw the cap back onto the body. Inspect the membrane for tears or leakage. The membrane should be uniformly stretched across the inner body.



Storage

Short term (up to one week)

1. Inspect the membrane for tears or leakage. The membrane should be uniformly stretched across the inner body
2. Replace the membrane if any damage has occurred.


Long term


1. Carefully unscrew the cap from the body.
2. Rinse the inside of the membrane cartridge with DI water.
3. Rinse the soaked portions thoroughly with DI water. Blot dry with tissue paper.
4. Screw the cap back onto the body without the electrolyte

Data logging and analysis

MiLAB™


1. Take your einstein™ Tablet or pair your einstein™ LabMate with your Android or iOS tablet via Bluetooth
2. Insert the sensor cable into one of the sensor ports
3. Launch MiLAB
4. MiLAB will automatically detect the sensor and show it in the Launcher View
5. Tap the Settings icon to set the sensor to measure dissolved Oxygen (DO) or atmospheric Oxygen (O₂)




6. Check the icon next to the sensor () to enable it for logging

MiLAB™ Desktop

1. Pair your einstein™ LabMate with your PC, MAC, or Linux machine via Bluetooth, or connect it via the USB cable (found in the einstein™ LabMate box).
2. Insert the sensor cable into one of the sensor ports.
3. Launch MiLAB.
4. MiLAB will automatically detect the sensor and show it in the Current Setup Summary window
5. Click Full Setup, located at the bottom of the Current Setup Summary window to program the data logger's sample rate, number of samples, units of measurement, and other options.



6. Click the Run button () on the main toolbar of the Launcher View to start logging.

Experiment set up

The Oxygen Sensor comes with:



- One Oxygen sensor
- One electrode
- One bottle Galvanic O₂ Fill Solution
- One bottle Salt for zero solution – (Sodium Sulfite, Na₂SO₃)
- One plastic dropper
- Extra membranes

Calibration

When calibrating the Oxygen sensor, you will have the choice of calibrating for Dissolved Oxygen (DO) or gaseous oxygen (O₂).


DO Calibration

MiLAB™ Desktop Set Zero Calibration

In the Current Reading column, click Set  to set the current value as the zero or base value. Reset  cancels this action

MiLAB™ Set Zero Calibration


Under Current Reading column flip the Set As Zero switch to set the current value as the zero or base value.




Calibrating in MiLAB™ Desktop



Two Point Calibration

1. Go to the Full Setup window and in the Calibrate column click Set



2. The Calibration window will appear




3. Saturate water with Oxygen by vigorously stirring the water. Enter the known value for saturated water according to your location in the Point 1, Real Reading field
4. Measure the solution and wait for the readings to stabilize. Enter the measurement as the Point 1 Measured Reading field and click the lock button 
5. Prepare a second solution with a known value (we recommend using the salt for zero solution in order to reduce the DO of water to zero). Enter this known value in the Point 2, Real Reading field
6. Measure the solution and wait for the readings to stabilize. Enter the measured value as the Real Reading on the first line and click the lock button 
7. Click Calibrate

Calibrating in MiLAB


Note: Graphics in your version of MiLAB may differ



Two Point Calibration

1. Tap the Settings button next to the sensor's name



2. Flip the Calibrate switch to On




3. Saturate water with Oxygen by vigorously stirring the water. Enter the known value for saturated water according to your location in the Point 1, Real Reading field
4. Measure the substance and wait for the readings to stabilize. Enter the value as the Point 1 Measured Reading field and click the lock button 
5. Prepare a second solution with a known value (we recommend using the salt for zero solution in order to reduce the DO of water to zero). Enter this known value in the Point 2, Real Reading field
6. Measure the solution and wait for the readings to stabilize. Enter the known value as the Real Reading on the first line and click the lock button 
7. Click Calibrate

O₂ Calibration


Calibrating in MiLAB™


One Point Calibration

1. Tap the Settings button next to the sensor's name



2. Flip the Calibrate switch to On




3. Choose a well-ventilated location (preferably outside)
4. Enter 20.9 (the percentage of oxygen in the air) in the Point 1, Real Reading field
5. Hold the electrode in the air and wait for the readings in the Current Reading field to stabilize.
6. Enter the Current Reading value as the Measured Reading field and click the lock button 
7. Click Calibrate.


Calibrating in MiLAB™ Desktop


One Point Calibration

1. Choose a well-ventilated location (preferably outside)
2. Go to the Full Setup window and in the Calibrate column click Set



3. The Calibration window will appear



4. Enter 20.9 (the percentage of oxygen in the air) in the Point 1, Real Reading field
5. Hold the electrode in the air and wait for the readings in the Measured Reading field to stabilize and click the lock button .
6. Click Calibrate

Troubleshooting

If the Oxygen sensor isn't automatically recognized by MiLAB, please contact Fourier Education's technical support.

Technical support

For technical support, you can contact the Fourier Education's technical support team at:
 Web: www.einsteinworld.com/support

Email: support@fourieredu.com

Copyright and Warranty

All standard Fourier Systems sensors carry a one (1) year warranty, which states that for a period of twelve months after the date of delivery to you, it will be substantially free from significant defects in materials and workmanship.

This warranty does not cover breakage of the product caused by misuse or abuse.

This warranty does not cover Fourier Systems consumables such as electrodes, batteries, EKG stickers, cuvettes and storage solutions or buffers.

©Fourier Systems Ltd. All rights reserved. Fourier Systems Ltd. logos and all other Fourier product or service names are registered trademarks or trademarks of Fourier Systems. All other registered trademarks or trademarks belong to their respective companies.

ALBERT EINSTEIN and EINSTEIN are either trademarks or registered trademarks of The Hebrew University of Jerusalem. Represented exclusively by Green Light. Official licensed merchandise. Website: einstein.biz